

Declaration of Conformity

We,

Tobin Sweden AB Arntorpsgatan 26 SE-442 45 Kungälv Sweden

| Art.Nr | DEVICE | Medical |
|--------|---|--------------|
| | | Device Class |
| 78 | TOBIN Eyewash Saline Solution 1000 ml (Wero REF 280029) | Is |
| 77 | TOBIN Eyewash Phosphate Buffered Solution 1000 ml (Wero REF 280048) | Is |
| 101 | TOBIN Eyewash Borate Buffered Solution 1000 ml | Is |
| 79 | TOBIN Pocket Eyewash Saline Solution 200 ml (Wero REF 280028) | Is |
| 94 | TOBIN Pocket Eyewash Phosphate Buffered Solution 200 ml (Wero REF 280044) |) Is |
| 102 | TOBIN Pocket Eyewash Borate Buffered Solution 200 ml | Is |
| 92 | TOBIN Pocket Eyewash Acetic Acid Solution 200 ml | Is |
| 125 | TOBIN Eyewash Saline Solution 500 ml (Wero REF 280069) | Is |
| 427 | TOBIN Eyewash Borate Buffered solution 500 ml | Is |
| 227 | TOBIN Eyewash Phosphate Buffered Solution 500 ml (Wero REF 280063) | Is |
| 120 | TOBIN Eyewash Saline Solution 20 ml (Wero REF 280050) | Is |

hereby declare that the above devices in Tobin Eyewash Systems, Class I(s), comply with applicable requirements in the Swedish Act **2021:600** concerning medical devices as well as Swedish Medical Products Agency regulations **LVFS 2003:11** on medical devices. The products also meet the requirements in European Directive **93/42/EEC.**

We comply with the applicable requirements in the **European Standard EN 15154-4:2009 Emergency** safety showers - Part 4: Non plumbed-in eyewash units.

Medical devices verified by our Notified Body RISE # 0402

EC-certificate and ISO 13485 certificate are available upon request.

Kungälv, 2021-08-18

Tobin Sweden AB

Stig Andersson

Managing director