

# Slishman Traction Splint - Compact (STS-C)

## INSTRUCTIONS FOR USE



1. Press silver spring button to extend outer tube for patients taller than 5ft (1.5m).



Secure outer tube using spring button snapped into either of two outer tube holes.



2. Apply neoprene outer tube strap firmly to proximal calf or patella position.



3. Apply groin strap, adjusting until snug around the upper thigh.



4. Gently pull cord for traction.



5. Once pain is controlled, lock cord in place by pulling down into V notch at top of inner tube.



6. Secure loose end of traction cord by wrapping around both cleat wings.



7. Stabilize\* and pad the injured limb for comfort.



8. Monitor CSM (circulation, sensation and motor function) closely and adjust as needed.

\*Stabilization can be accomplished with many items including pillows, blankets, backboard straps, vacuum splints, malleable splints, or box splints. The Slishman Pressure Wrap is included to help with rotational stability using the following steps:

1. Lasso either foot with the loop of the elastic wrap. For a very small foot you may choose to cinch the elastic to be snug, but the loop does not need to be tight.
2. Wrap around the opposite foot such that the two feet are adjacent, thus providing rotational stability. As you wrap, fasten the white hooks to the black fabric.
3. Continue to wrap to the end of the elastic and secure with the white hook fabric under the "Slishman Pressure Wrap" label to the black fabric or use the provided safety pin.
4. Write the application time in the space provided on the label if deemed useful.  
(Wrap may also be used to limit bleeding as with any elastic pressure bandage)



# Slishman Traction Splint - Compact (STS-C)

## CLEANING AND DISPOSAL INSTRUCTIONS

RESCUE ESSENTIALS

The Slishman Traction Splint should be thoroughly cleaned by washing with warm water and disinfectant and then wiped down with a clean rag.

If the entire device is contaminated,\* place it in a red biohazard bag and dispose according to your hazardous waste protocol.

Contaminated hip/groin strap can be removed from the device and laundered in the washing machine with detergent and disinfectant added. Do not use bleach.

Remove the strap and hang to dry. Do not put straps in dryer.

The outer tube strap is neoprene and should not be cleaned using hot water. Instead, wipe the strap thoroughly using disinfecting wipes.

The straps should be considered disposable if they are grossly contaminated or if cleaning requires exuberant effort. These items should be replaced rather than cleaned.

\*If the splint comes in contact with toxins like organophosphates, radioactive material, or excessive bodily fluids, then it should be disposed of as deemed appropriate for hazardous materials or biological wastes. If it can be washed and sanitized, then reuse is acceptable.

## STS-C COMPONENTS



RESCUE ESSENTIALS

[www.rescue-essentials.com](http://www.rescue-essentials.com)

TRITECH FORENSICS  
*Identify. collect. preserve.*

Manufactured by Tri-Tech Forensics, Inc.  
dba Rescue Essentials under U.S. Patents: 6,394,972 -  
6,786,882 - 6,913,587. Additional Patent Pending

REV. 9/3/19

910-830-0286 [info@rescue-essentials.com](mailto:info@rescue-essentials.com)



8770 Trade St Leland, NC 28451

CE EC REP

Mdi Europa GmbH  
Langenhagener Str. 71  
D-30855 Langenhagen

MD NON STERILE LATEX